


# GuReater™

CE<sub>2292</sub>

## CoCr Sirolimus-Eluting Coronary Stent System



## Product Information

### Features

- CoCr drug-eluting stent
- 75µm strut thickness
- Low recoil ratio
- PLGA biodegradable polymer
- Effective drug of Sirolimus
- 90-day drug releasing period
- MRI compatible

### Benefits

- Super flexibility and crossability
- Excellent compliance
- Great post-stenting performance
- Excellent drug releasing control
- Scientific drug releasing period
- Avoid of inflammation caused by durable polymer
- Earlier endothelialization

### Ranges



- Ø 2.5 - 4.0 mm
- 12 - 36 mm multi stent lengths

### PLGA Biodegradable Polymer of Drug Coating

Drug completely released in 90 days  
PLGA coating completely degraded in 180 days

### 75µm CoCr Strut

Provide super flexibility and crossability  
Ensure the great compliance after implantation



Microscopic view of stable drug coating for GuReater™ during degradation

# Technical Parameters

Stent Material	L605 CoCr Alloy
Polymer Material	PLGA
Strut Thickness	75 $\mu$ m
Metal-artery Ratio	15%
Recoil Ratio	3.4%
Shorting Ratio	< 1%
Lowest Crossing Profile	0.035" (0.90mm)
Balloon Tip Profile	0.017"
Delivery System	RX

# Ordering Information

Stent Length (mm)	Stent Diameter $\varnothing$ (mm)				
	2.5	2.75	3.0	3.5	4.0
12	LPRPC2512	LPRPC2712	LPRPC3012	LPRPC3512	LPRPC4012
15	LPRPC2515	LPRPC2715	LPRPC3015	LPRPC3515	LPRPC4015
18	LPRPC2518	LPRPC2718	LPRPC3018	LPRPC3518	LPRPC4018
21	LPRPC2521	LPRPC2721	LPRPC3021	LPRPC3521	LPRPC4021
24	LPRPC2524	LPRPC2724	LPRPC3024	LPRPC3524	LPRPC4024
29	LPRPC2529	LPRPC2729	LPRPC3029	LPRPC3529	LPRPC4029
33	LPRPC2533	LPRPC2733	LPRPC3033	LPRPC3533	LPRPC4033
36	LPRPC2536	LPRPC2736	LPRPC3036	LPRPC3536	LPRPC4036

Official Partner **MEDITRANS** 



**Jakarta** : Jl Munggang no. 86-87 Balekambang  
Kramat Jati, Jakarta Timur, 13530  
**Surabaya** : Rungkut Menanggal Harapan  
Blok O No. 28, Surabaya 60293  
**Others** : Bandung, Makassar, Medan, Balikpapan



(+62) 21 2204 7248  
(+62) 31 8786 2593  
(+62) 821 8000 1102



[www.meditrans.id](http://www.meditrans.id)



[info@meditrans.id](mailto:info@meditrans.id)



[meditrans.id](https://www.facebook.com/meditrans.id)



[meditransid](https://www.instagram.com/meditransid)